



BLUEFAN.NL

Custom Lasers - Optics - Power meters - Laser Diode Drivers - Advanced Test Loads
Lab Service - Measurement - Calibration - Adjustment - Repair - Modification



Calibration Certificate

Customer:

Name:

Bluefan.nl

Email:

coherentwave@gmail.com

Laser Power Meter

Description of LPM

Bluefan Basic LPM, serial no. 11-50-0001

Accuracy:

5%

Spectral range:

400nm-1100nm

Warm-up time

30 minutes

Adjustment:

Yes

Recoating:

No

Measurements:

Measurement range

2W

Resolution

1mW

Wavelength

447nm

Applied Power:

1.0000W

Measured power:

1.000W

981404:44:43.00

Room temperature:

21 C

Date:

14. December 2011

Measurements are done as accurate as possible using a Laser Precision Rk-5710 with a RkP-575 probe. Bluefan.nl cannot be held responsible for any incorrect measurements and its consequences. Bluefan.nl can be contacted at coherentwave@gmail.com